

**TEMPERATURE CALIBRATION CERTIFICATE**

This document may not be copied without written authorisation of Templog

Certificate No:	1510010012	Certification Date:	09.02.2018		
Calibration ordered by:	Templog	Calibration Performed by	Semih Tan		
		Calibration Checked by	Sezer Alagöz		
Temperature:	20 °C	Humidity:	40% RH ±5	Pressure:	990±5mbar
Product code	TemplogONE	Seial Number	1510010012	New or Used	New

**Test Equipment used**

Manufacturer	LUTRON		Device Name	YK-90HT	
Traceability	UML	Serial	Q850919	Calibration Date	14.11.2017

**CALIBRATION TEST DETAILS**

Serial Number:	1510010012	Temperature Range:	-40°C to +70°C
Display Resolution:	0.25 °C	Accuracy :	±0.1 °C

Reference Equipment °C	Device Calibrated °C	Deviation °C	Uncertainty ±°C	PASS/FAIL
-10.0	-9.75	-0.25	0.5	PASS
0.0	0.25	-0.25	0.5	PASS
10.3	10.00	0.30	0.5	PASS
20.0	19.75	0.25	0.5	PASS

Method of Testing: Details of testing method is available on request

Test Result PASS-FAIL	PASS	Signature of Calibration Technician	
-----------------------	------	-------------------------------------	---